


VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska


Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program


Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130526
A	Name	Joseph V. Schneider	ID# 4188 Date 8-25-11
	Agency	SDPI (SAND POINT P.D.)	Phone # 907-383-3700
	Instrument Location	SAND POINT, ALASKA 99661	
B	Alco S/N	3B	Target Value .080 High Pressure 800
	Alco Test Value Average	.076	.076
		1 st Alco	2 nd Alco
	Signature		
(OVER)			

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 2 day of Sept, 2011. 


Nikki Roth, Notary Public
My Commission Expires With Office

BT9 03/00



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130526

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130526

AUGUST 25, 2011

OPERATOR'S NAME:

SHOENAKER J V

OPERATOR'S NUMBER: 4188

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

L

O.L. #: U

DEPT/AGENCY: SDPI

CASE/REPORT: J

TEST TYPE: V

ALCO TARGET VALUE: .000

ALCO S/N: 3B

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 29.86 in
ALCO TARGET .079 23:05
BLANK TEST .000 23:06
INTERNAL STANDARD VERIFIED 23:06
ALCO TV 29.86 in .076 23:07
BLANK TEST .000 23:08
SUBJECT SAMPLE .000 23:08
BLANK TEST .000 23:09
ALCO TV 29.86 in .076 23:09
BLANK TEST .000 23:10

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130526

AUGUST 25, 2011

TIME 23:11

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 43c
BREATH TUBE: 41c
BAROMETER: 29.86 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!#\$%&'()*+,-./0123456789:;<=>?@ABCDEF
GHIJKLMNOPQRSTUVWXYZ[\]^_`abcde{ghijklmno
pqrstuvwxyz{|}~